Blanchard Valley Health System Site:	Loca	ation:			
Training / Competency As sessment for Laboratory Point of Care Testing			Date of Hire / Transfer:		
Method: POCT DCA Vantage Analyzer		Date of Ini	tial Training:		
Test Analyte: Hemoglobin A1c (HbA1c)	Supervis	sor:			
Associate Name and Title:					
(print legibly) Associate ID number:					
CLIA Complexity: Waived (2 areas must be as	sessed) (CAPappro	ved areas 1 – 6) (JC	approveda	reas 1, 3, 5, 6)	
Training assessment after initial training (ne	w hires) and/or new	method and/or new te	est analytes	Date:	
Competency assessment annually			Date Due:		
1 2 3 Perform QC testing a acceptable test results meets CAP/JC comp				•	
HbA1c : QC Normal Control	c: QC Normal Control QC Abnormal Control			PASS / FAIL (circle)	
4 Perform optical test check as require maintenance and function checks under dire com petency requirements for element 4.) Re	ect observation and	· · ·		0	
Mean transmittance (Mean T) 0.95 – 1.05		ACCEPTABLE?	YES/NO	(circle)	
Standard Deviation (SD T) <0.0015		ACCEPTABLE?	YES / NO	(circle)	
Drift T No applicable limits		ACCEPTABLE?	YES/NO	(circle)	
Drift Max <0.0280		ACCEPTABLE?	YES/NO	(circle)	
				PASS / FAIL (circle)	
5 Perform testing on an unknown sam results meets CAP/JC competency requirem			nsamplean	d getting acceptable test	
UNKNOWN Test: (Patient or Control)	HbA1c			PASS / FAIL (cirde)	
Associate Signature:			Date:		
Signature of Qualified Trainer/Assessor:			_ Date:		
Job title of Qualified Trainer/Assessor:			1-5 Satisfactory / Unsatisfactory		
6An evaluation of problem solving ski meet CAP/JC competency requirements for	lsbywrittenexam element6.	in Net Learning with	70% pas sir	ng score will be assigned to	
Assigned: Due:	Completed:	Score	9:	PASS / FAIL (circle)	
		Reviewed by:			
		· -		Date:	