

Blanchard Valley Health System: Point of Care Testing

ID NOW™ External Quality Control Log

External QC testing is recommended:

- When a new shipment of kits is received
- When a new untrained operator performs testing
- The first time an assay is run on an instrument
- When required by local, state, and/or federal regulations, accrediting groups, or your lab's Quality Control procedures

| DATE | ASSAY (CIRCLE TYPE) | INSTRUMENT SN | ID NOW™ KIT LOT/EXP | POSITIVE CONTROL LOT/EXP | NEGATIVE CONTROL LOT/EXP | POSITIVE CONTROL RESULT | NEGATIVE CONTROL RESULT | TESTER'S INITIALS | COMMENTS |
|------|--|---------------|---------------------------|--------------------------------|--------------------------------|-------------------------------|-------------------------------|----------------------|----------|
| | Flu Strep A RSV COVID-19 2.0* | | | | | | | | |
| | Flu Strep A RSV COVID-19 2.0 | | | | | | | | |
| | Flu Strep A RSV COVID-19 2.0 | | | | | | | | |
| | Flu Strep A RSV COVID-19 2.0 | | | | | | | | |
| | Flu Strep A RSV COVID-19 2.0 | | | | | | | | |
| | Flu Strep A RSV COVID-19 2.0 | | | | | | | | |
| | Flu Strep A RSV COVID-19 2.0 | | | | | | | | |

REVIEWED BY: _____

DATE: _____