

Year: _____

**Blanchard Valley Health System: Point of Care
Refrigerator Temperature Log**

January 2-8 C	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	25	26	27	28	29	30	31	
Temp:																																
February 2-8 C	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	25	26	27	28	29	30	31	
Temp:																																
March 2-8 C	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	25	26	27	28	29	30	31	
Temp:																																
April 2-8 C	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	25	26	27	28	29	30	31	
Temp:																																
May 2-8 C	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	25	26	27	28	29	30	31	
Temp:																																
June 2-8 C	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	25	26	27	28	29	30	31	
Temp:																																

Supervisory Review: _____ Date: _____

If temp out of range, move testing supplies to area within temp range. Then adjust out of range temp and recheck in 4 hrs.
If temp continues out of range, after the recheck, initiate a maintenance work order. Document date and time work order sent below or on the back of this form (include dates of any actions taken). All corrective actions must be documented. Be sure to document when corrective actions are completed and reagents moved back to original storage.

Year: _____

**Blanchard Valley Health System: Point of Care
Refrigerator Temperature Log**

July 2-8 C	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	25	26	27	28	29	30	31	
Temp:																																
August 2-8 C	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	25	26	27	28	29	30	31	
Temp:																																
September 2-8 C	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	25	26	27	28	29	30	31	
Temp:																																
October 2-8 C	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	25	26	27	28	29	30	31	
Temp:																																
November 2-8 C	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	25	26	27	28	29	30	31	
Temp:																																
December 2-8 C	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	25	26	27	28	29	30	31	
Temp:																																

Supervisory Review: _____ Date: _____

If temp out of range, move testing supplies to area within temp range. Then adjust out of range temp and recheck in 4 hrs.
 If temp continues out of range, after the recheck, initiate a maintenance work order. Document date and time work order sent below or on the back of this form (include dates of any actions taken). All corrective actions must be documented. Be sure to document when corrective actions are completed and reagents moved back to original storage.